

Laboratório	A1	A2	A11	A23	A24	A25	A32	A36	A68	A69	A80	B13	B18	B27	B37	B38	B39	B44	B45	B47	B49	B51	B52	B53	B54	B55	B56	B57	B58	B59	B60	B61	B63	B64	B65	B67	B78	
84	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	
90	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	
110	Green	Green	Green	Green	Green	Green	Green	Red	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	
Laboratório	A1	A2	A11	A23	A24	A25	A32	A36	A68	A69	A80	B13	B18	B27	B37	B38	B39	B44	B45	B47	B49	B51	B52	B53	B54	B55	B56	B57	B58	B59	B60	B61	B63	B64	B65	B67	B78	
84	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	
90	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green
110	Green	Green	Green	Green	Green	Green	Green	Red	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	Green	
Especificidades reportadas	A2	A23	A24	A25	A32	A68	A69	B27	B44	B49	B51	B52	B57	B58	B63	A1	A11	A80	B13	B37	B47	B53	B59	A36	B18	B38	B39	B45	B54	B55	B56	B60	B61	B64	B65	B67	B78	
Nº de laboratórios	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	2	2	2	2	2	2	2	2	2	1	1	1	1	1	1	1	1	1	1	1	1		
%	100%															67%							33%															
Amostra sem consenso: Somente três laboratórios reportaram seus resultados. É exigido para consenso um mínimo de dez laboratórios para validação da amostra conforme item 3 da carta de Instruções.																Especificidades intermediárias reportadas pelos laboratórios participantes com 67%.							Especificidade reportada por um único laboratório participante.															